

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

MSP - GLOBAL, s.r.o.,
CAB number 2286, Czech Calibration Service
Dobrá 240, 739 51 Dobrá

CMC for the field of measured quantity: Mass

Ordinal number ¹	Calibrated quantity / Subject of calibration	Nominal range			Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work-place					
		min.	unit	max.						unit				
1*	Electronic and mechanical scales with non-automatic function	0 kg		up to	0.22 kg		class F1	weight						
		0.22 kg		up to	1.6 kg		class F1	weight						
		1.6 kg		up to	5 kg		class F1	weight						
		0.5 kg		up to	10 kg		class F1	weight						
		10 kg		up to	16 kg		class F2	weight						
		16 kg		up to	20 kg		class F2	weight						
		20 kg		up to	50 kg		class M1	weight						
		50 kg		up to	100 kg		class M1	weight						

¹ Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

² The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.

³ If the document identifying the test procedure is dated, only these specific procedures are used. If the document identifying the test procedure is not dated, the latest edition of the specified procedure is used (including any changes).

